



Certificate of Analysis

1.08418.0000 Titriplex® III for analysis (ethylenedinitrilotetraacetic acid, disodium salt dihydrate) ACS,ISO,Reag. Ph Eur
Batch K56581818

	Spec. Values		Batch Values	
Assay (complexometric; calculated as dihydrate)	99.0 - 101.0	%	100.1	%
Identity (IR-spectrum)	passes test		passes test	
Identity (Complexation)	passes test		passes test	
Identity (Sodium)	passes test		passes test	
Appearance	passes test		passes test	
Appearance of solution (50 g/l; water)	passes test		passes test	
pH (50 g/l; water)	4.0 - 5.0		4.5	
In water insoluble matter	≤ 0.003	%	≤ 0.003	%
Chloride (Cl)	≤ 0.004	%	≤ 0.004	%
Sulfate (SO ₄)	≤ 0.01	%	≤ 0.01	%
Cyanide (CN)	≤ 0.001	%	≤ 0.001	%
Heavy metals (as Pb)	≤ 0.0005	%	≤ 0.0005	%
Ca (Calcium)	passes test		passes test	
Cu (Copper)	≤ 0.0001	%	≤ 0.0001	%
Fe (Iron)	≤ 0.0005	%	≤ 0.0005	%
Pb (Lead)	≤ 0.001	%	≤ 0.001	%
Nitrilotriacetic acid (HPLC)	≤ 0.05	%	0.02	%
Loss on drying (150 °C; 6 h)	8.7 - 11.4	%	9.7	%

Date of release (DD.MM.YYYY) 18.03.2025
Minimum shelf life (DD.MM.YYYY) 31.03.2030

Dr. Sebastian Lips
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.